

Quick Reference Instructions

Win against asthma with help

from 



BEFORE USING, READ ENTIRE PARI VORTEX® INSTRUCTIONS FOR USE FOR SAFETY PRECAUTIONS, CLEANING METHODS, AND PATIENT INFORMATION.

1

Look for foreign, damaged, and/or missing parts. If using an optional mask, firmly attach to mouthpiece.

4

Use with Mouthpiece: Place mouthpiece into mouth. Exhale, depress MDI, and then breathe in slowly and deeply. Hold breath for 10 seconds or for as long as is comfortable.

2

Remove cap of metered dose inhaler (MDI) and insert into the back MDI receptacle of the PARI VORTEX® as shown.

5

Use with Mask: Apply to face over nose and mouth creating a good seal. Depress MDI and breathe in and out slowly for 10 seconds.

3

Shake MDI per instructions.

6

Follow physician and MDI instructions before repeating steps 4-5.

Request the PARI VORTEX® from your pharmacist!

VORTEX[®]

Non Electrostatic Valved Holding Chamber

SINGLE PATIENT USE

LATEX FREE • DISHWASHER SAFE



IMPORTANT

Follow instructions included with your metered dose inhaler (MDI).

Before using, read entire PARI VORTEX[®] instructions for use for safety precautions, cleaning methods, and patient information.



CAUTION

U.S. Federal Law restricts this device to sale by or on the order of a physician. Use only when prescribed by a physician and with the prescribed medication. Read all warnings and instructions before use. Use the product only for its intended use as described in these Instruction For Use. Improper use can result in serious or fatal illness/injury.



WARNING

1. Close supervision is necessary when this product is used by children or near physically or mentally impaired. This device contains small parts that may present a choking hazard to small children. Always keep out of the reach of children three years old and under. All parts should be seated firmly in place. Use of an improperly assembled PARI VORTEX[®] Non Electrostatic Valved Holding Chamber could prevent adequate delivery of medication. If not sure of the suitability and/or functioning of your MDI, consult MDI instructions and/or your physician.
2. To reduce the risk of infection, illness or injury from contamination and to reduce the risk of improper performance and treatment, you must thoroughly clean and dry all parts of the PARI VORTEX[®] as instructed in the PARI VORTEX[®] instructions.
3. To reduce the risk of improper treatment, use only PARI accessories.
4. To reduce the risk of improper treatment, product should be replaced if valves or parts are damaged or missing.
5. To reduce the risk of improper treatment, make sure duckbill valve is properly fitted.
6. If unsure on how to use this product, talk to your physician or pharmacist.
7. Administer medication per MDI instructions.



Developed by:

PARI Respiratory Equipment, Inc.

2412 PARI Way, Midlothian, VA 23112 USA

Tel: 800.FAST.NEB (327.8632) and 804.253.7274

Web: **PARI**.com

US Patent No. 7,562,656

© 2015 PARI Respiratory Equipment, Inc.

051D0023 Rev C 9-15